

A

***Neisseria gonorrhoeae* (NCTC 8375 [B5025] Strain) Bacterial qDNA**

Catalog Number: 17-924-500

Lot Number: G0302

Product Description: *Neisseria gonorrhoeae* (NCTC 8375 [B5025] Strain) Bacterial qDNA PCR control

Unit Size: 500 µL

Expiration Date: 03-2017

Suspending Buffer: 10 mM Tris, 1 mM EDTA, pH 8.0

QUALITY CONTROL DATA

DNA Copy Number by Quantitative Realtime PCR: 5.6 x 10⁵ DNA copies per milliliter

PCR Analysis of DNA Control: PCR analysis was performed on purified bacterial DNA using primers specific for the B protein gene of *Neisseria gonorrhoeae*. The reaction produced a 391 bp fragment. The band was observed from 10⁵ copies/reaction down to 10³ copies/reaction.

DNA Quantitation: DNA copy number was determined by quantitative realtime PCR. DNA copy number may vary depending on the quantitation method and primers used.

PRODUCT DETAILS

Shipping and Storage: This product is shipped frozen on dry ice. **Store at -20°C upon receipt.** Avoid multiple freeze-thaw cycles as product degradation may result.

Recommendations: Upon thawing, centrifuge the vial for a few seconds to remove residual droplets from the lid. **CAUTION:** ABI does not recommend storage of dilutions of quantitated PCR controls under any conditions. All dilutions should be made immediately before use and used promptly. We have observed that dilutions used for standard curves may tend to "lose" copy number with time (sometimes a matter of an hour or so after dilution).

Applications For Use: ABI's quantitated PCR controls are prepared from virus, bacteria, parasites, or mollicutes, and are intended for use as positive PCR quantitation standards for the organism in question. Due to the nature of these products, ABI cannot guarantee their suitability as extraction controls. Additionally, due to the extreme sensitivity of detection in PCR reaction, and since no method of purification can guarantee the complete absence of extraneous agents, PCR controls are not intended for use as negative controls for other organisms.

Safe Handling Recommendation: The DNA extraction procedure used has been shown to eliminate the infectivity of most viruses and bacteria; therefore, this product is not considered biohazardous. However, this product is not specifically tested and should be handled in accordance with Good Laboratory Practices and any applicable local guidelines.

**This product is for research use only.
Not for use in diagnostic procedures.**



Quality Control

03-14-2016

Date